

**Audiometric Calibration**  
**The Audiology Center**  
**1364 Ebenezer Road**  
**Rock Hill, SC 29732**

Frequency Dial (Hz)	Air Conduction Level At 70 dB HL Per ANSI 1969 (dB SPL)	Acceptable Range Per ANSI 1969 (dB SPL)	Actual Air Conduction Level (dB SLP)	
			CH 1	CH 2
			Right	Left
125	114.8	109.8 - 119.8		
250	95.5	92.5 - 98.5		
500	81.9	78.9 - 84.9	80	80
750	78.5	75.5 - 81.5		
1000	77.1	74.1 - 80.1	80	80
1500	76.4	73.4 - 79.4		
2000	78.8	75.8 - 81.8	79	79
3000	78.4	75.4 - 81.4		
4000	75.5	71.5 - 79.5	78	78
6000	78.8	73.8 - 83.8		
8000	79.4	74.4 - 84.4		

Frequency Dial (Hz)	Acceptable Frequency Range Per ANSI 1969 (Hz)	Actual Frequency	
		CH 1	CH 2
		Right	Left
125	121 - 129		
250	242 - 258		
500	485 - 515	500	500
750	727 - 773		
1000	970 - 1030	1000	1000
1500	1455 - 1545		
2000	1940 - 2060	2000	2000
3000	2910 - 3090		
4000	3880 - 4120	4000	4000
6000	5820 - 6180	6000	6000
8000	7760 - 8240	8000	8000

Technician: OHM User

Date: 12/31/04

Company: OHM Demo Company

AUDIOMETER

Model Number: MI-5000

Serial #: 20604

Manufacturer: Monitor

Earphone Type: TDH 49

Cushion Type: MX-41/AR